

**QUALITY CONTROL CERTIFICATE**

**TOBRAMYCIN 10 µg Neo-Sensitabs**

REF: 62112

LOT: 24-1018

Date of analysis: 2024-08

Date of expiry: 2028-08

Appearance: Light yellow, plane tablet, 9 mm in diameter, on both sides printed with a black identification code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:  
*E. coli* ATCC 25922  
*Staphylococcus aureus* ATCC 29213  
*Pseudomonas aeruginosa* ATCC 27853

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA  
 Chief Medical Officer, Head of Quality