

QUALITY CONTROL CERTIFICATE

AMOXYCILLIN+CLAVULANATE 2+1 µg Neo-Sensitabs

REF: 69812

LOT: 24-1010

Date of analysis: 2024-08

Date of expiry: 2026-07

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E.coli ATCC 35218

Haemophilus influenza ATCC49766

Staphylococcus aureus ATCC 29213

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality