

QUALITY CONTROL CERTIFICATE

**AMPICILLIN 2 µg Neo-Sensitabs**

REF: 66712

LOT: 24-1021

Date of analysis: 2024-09

Date of expiry: 2029-09

Appearance: Light yellow, plane tablet, 9 mm in diameter, on both sides printed with a black identification code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*Ent. Faecalis* ATCC29212  
*Strep. Pneumonia* ATCC49619  
*Haemophilus influenza* ATCC49766  
*Staph.Aureus* ATCC29213

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostics ApS



Claus Bo Svendsen MD PhD MBA  
Chief Medical Officer, Head of Quality