

## QUALITY CONTROL CERTIFICATE

### **PENICILLIN 1 U Neo-Sensitabs**

REF: 66912

LOT: 24-1022

Date of analysis: 2024-09

Date of expiry: 2028-09

Appearance: Light yellow, plane tablet, 9 mm in diameter, on both sides printed with a black identification code.  
Complies.

Thickness: Mean of 1.50-1.65 mm.  
Complies.

### Performance:

Test strains:

*Staphylococcus aureus* ATCC 29213  
*Streptococcus pneumoniae* ATCC 49619

*Product complies with specifications.*

Uniformity: Inhibition zones of individual tables vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies

Rosco Diagnostica ApS

A handwritten signature in blue ink that reads "Claus Bo Svendsen".

Claus Bo Svendsen MD PhD MBA  
Chief Medical Officer, Head of Quality