

## QUALITY CONTROL CERTIFICATE

### **AmpC Confirm ID Pack, Neo-Sensitabs**

REF: 98007

LOT: 24-4010

Date of expiry: 2025-06

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10 :1.50-1.65 mm.  
Complies.

Performance:

Test strains:	Results
E. coli ATCC 25922	Negative
Ent. Cloacae NCTC 13406	Amp C Positive
Ent. Cloacae ATCC BAA - 1143	Amp C Positive

Product complies with specifications..

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA  
Chief Medical Officer, Head of Quality