

QUALITY CONTROL CERTIFICATE**Total ESBL Confirmation Kit**

REF: 98014

LOT: 24-4011

Date of expiry: 2025-08

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains	:Results
E. coli ATCC 25922	Negative
K. pneumonia ATCC 700603	ESBL Positive
ENT. Cloacae NCTC 13464	ESBL Positive

Product complies with specification.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA  
Chief Medical Officer, Head of Quality